

September 20-21, 2018
Lisbon, PortugalMonika Kwacz et al., J Otol Rhinol 2018 Volume: 7
DOI: 10.4172/2324-8785-C2-011

NEW IMPLANTABLE CHAMBER STAPES PROSTHESIS: ASSESSMENT OF EFFECTIVENESS AND SAFETY

Monika Kwacz¹, Kazimierz Niemczyk² and Tadeusz Kwacz³¹Institute of Micromechanics and Photonics, Warsaw University of Technology, Poland²Medical University of Warsaw, Poland³Biomechatronics Ltd. in organization, Poland

New chamber stapes prosthesis (ChSP) is a medical device intended for patients with stapes otosclerosis. The ChSP (Fig.1A) consists of a chamber a membrane and a plate. The introduction of the ChSP to clinical practice requires a number of preclinical and clinical tests to assess its effectiveness and safety. First, the functioning of the ChSP was verified by numerical simulations. The results showed that the ChSP has a significant advantage over the piston prosthesis and effectively transmits vibration for frequencies 0.4-10 kHz. Next, the ChSP prototype (Fig.1B) was made and the experimental studies were carried out. The prototype was implanted in a human cadaveric temporal bone (Fig.1C, D). The vibrations of the round window membrane (RWM) were measured using a scanning laser Doppler vibrometer before and after experimental stapedotomy. The vibrations were induced by the acoustic stimulation (AC). The sound (90 dB SPL, 0.8-8 kHz) was introduced to the external ear canal via a loudspeaker and controlled by a microphone. Based on the measured displacement amplitudes, the magnitude of fluid volume displacement at the RWM was calculated in both the normal and the implanted ear. The experiments confirmed that the ChSP prototype works correctly and effectively transmits sound for 0.8-8kHz. In subsequent experiments, we measured pressure in the perilymph fluid before and after ChSP-stapedotomy. The pressure was induced by the AC and measured using two optical pressure sensors placed in the scala vestibule and the scala tympani. The differential intracochlear pressure was calculated and the results were compared between the healthy and implanted ears. Finally, we performed a preliminary risk analysis and assessed the risk-benefit ratio according to ISO 14971

Biography

Monika Kwacz has completed her PhD from Warsaw University of Technology (Poland). She works as an Associate Professor at the Institute of Micromechanics and Photonics. She conducts biomedical research, manages R&D projects and is the Scientific Supervisor of the PhD students. She has published more than 25 papers in reputed journals.

m.kwacz@mchtr.pw.edu.pl

Any Comments:

Funding: This work was partially supported by the Polish National Centre for Research and Development (Research Project No. 266457/NCBR/2015). There was no additional external funding received for this study.




Fig.1-The ChSP prototype (A,B) and experimental setup (C,D).
B-temporal bone, S-speaker, M-microphone




Fig.2-Fluid volume displacement at the RWM in the normal and implanted ear